



CORFLEX®

Designed for life.

CORFLEXGLOBAL.COM | 800 426 7353

CoolTex™ Ag 13" Knee-O-Trakker with Hinge Application Instruction

Fitting and Application:

1. Determine side of kneecap (patella) that requires support, secure buttress on appropriate side and rethread straps as needed through elastic in popliteal area. (Note: buttress may be removed and placed on opposite side.)



2. Grasp tabs at wider end of stabilizer noting seam positioned to back of leg. Insert foot into sleeve and pull onto leg.



3. Your kneecap (patella) should be centered in the circular opening with buttress placed on rim of the kneecap (patella). Fasten the support straps at the thigh and calf for comfortable fit without inhibiting circulation.



4. To adjust the amount of patellar buttress correction, grasp both small straps above and below kneecap by hook tip and pull back onto loop portion of strap. Adjust small straps to desired tension or comfort level.



CORFLEX®

Designed for life.

CORFLEXGLOBAL.COM | 800 426 7353

CoolTex™ Ag 13" Knee-O-Trakker with Hinge Application Instruction

Fitting and Application:

1. Determine side of kneecap (patella) that requires support, secure buttress on appropriate side and rethread straps as needed through elastic in popliteal area. (Note: buttress may be removed and placed on opposite side.)



2. Grasp tabs at wider end of stabilizer noting seam positioned to back of leg. Insert foot into sleeve and pull onto leg.



3. Your kneecap (patella) should be centered in the circular opening with buttress placed on rim of the kneecap (patella). Fasten the support straps at the thigh and calf for comfortable fit without inhibiting circulation.



4. To adjust the amount of patellar buttress correction, grasp both small straps above and below kneecap by hook tip and pull back onto loop portion of strap. Adjust small straps to desired tension or comfort level.



Care Instructions

The CoolTex Ag 13" Knee-O-Trakker may be hand washed using warm water and detergent for delicate fabrics. Remove hinges prior to washing. Repeatedly rinse and line dry. Do not use bleach or place in a washer or dryer.

Caution

This product has been designed and tested based on single patient usage. Since it is a single use device it should never be used by multiple patients. If additional pain or symptoms occur while using this device, seek medical attention.

Indications

Patella tracking abnormalities, tendonitis, knee strains or sprains, sports injuries, post-op rehabilitation and arthritis.

Contraindications

Not for use on patients with circulatory or sensation issues. If swelling, pain, skin irritation, or an unusual reaction occurs, discontinue use and consult your medical professional.

Warranty

Products manufactured by Corflex are warranted to the original purchaser against defects in workmanship and material for the life of the product. The foregoing warranty shall not apply to normal wear and tear; or if the product has been subjected to misuse or misapplication the warranty will be voided at the discretion of Corflex Global.

To Reorder

Product Number	Size	Measurement		
		6" Above Mid-Pat	Mid-Pat	6" Below Mid-Pat
85-7272	XS	16"-18"	12"-13"	11"-12"
85-7273	S	17"-19"	13"-14"	12"-13"
85-7274	M	18"-20"	14"-15"	13"-14"
85-7275	L	19"-21"	15"-16"	14"-15"
85-7276	XL	20"-22"	16"-18"	15"-16"
85-7277	2XL	21"-23"	18"-20"	16"-17"
85-7278	3XL	23"-26"	20"-22"	17"-18"
85-7279	4XL	26"-29"	22"-24"	18"-19"

Care Instructions

The CoolTex Ag 13" Knee-O-Trakker may be hand washed using warm water and detergent for delicate fabrics. Remove hinges prior to washing. Repeatedly rinse and line dry. Do not use bleach or place in a washer or dryer.

Caution

This product has been designed and tested based on single patient usage. Since it is a single use device it should never be used by multiple patients. If additional pain or symptoms occur while using this device, seek medical attention.

Indications

Patella tracking abnormalities, tendonitis, knee strains or sprains, sports injuries, post-op rehabilitation and arthritis.

Contraindications

Not for use on patients with circulatory or sensation issues. If swelling, pain, skin irritation, or an unusual reaction occurs, discontinue use and consult your medical professional.

Warranty

Products manufactured by Corflex are warranted to the original purchaser against defects in workmanship and material for the life of the product. The foregoing warranty shall not apply to normal wear and tear; or if the product has been subjected to misuse or misapplication the warranty will be voided at the discretion of Corflex Global.

To Reorder

Product Number	Size	Measurement		
		6" Above Mid-Pat	Mid-Pat	6" Below Mid-Pat
85-7272	XS	16"-18"	12"-13"	11"-12"
85-7273	S	17"-19"	13"-14"	12"-13"
85-7274	M	18"-20"	14"-15"	13"-14"
85-7275	L	19"-21"	15"-16"	14"-15"
85-7276	XL	20"-22"	16"-18"	15"-16"
85-7277	2XL	21"-23"	18"-20"	16"-17"
85-7278	3XL	23"-26"	20"-22"	17"-18"
85-7279	4XL	26"-29"	22"-24"	18"-19"