

CORFLEX®


Designed for life.

CORFLEXGLOBAL.COM | 800 426 7353


CoolTex™ 13" Anterior Closure Knee Wrap with Hinge Application Instruction

Fitting and Application:


1. Open knee wrap and unfasten both hook and loop straps. Locate label on inside of brace. Place wrap behind the knee with label located at the top.





2. Center knee wrap on back of leg, position condyle pads at knee joint. Attach distal wrap closure snugly followed by proximal closure.



3. Fasten distal hook and loop strap to desired fit followed by proximal strap.



4. Adjust placement as required. Hinges are removable for washing.



CORFLEX®


Designed for life.

CORFLEXGLOBAL.COM | 800 426 7353


CoolTex™ 13" Anterior Closure Knee Wrap with Hinge Application Instruction

Fitting and Application:


1. Open knee wrap and unfasten both hook and loop straps. Locate label on inside of brace. Place wrap behind the knee with label located at the top.




2. Center knee wrap on back of leg, position condyle pads at knee joint. Attach distal wrap closure snugly followed by proximal closure.



3. Fasten distal hook and loop strap to desired fit followed by proximal strap.



4. Adjust placement as required. Hinges are removable for washing.



Care Instructions

The CoolTex Knee Anterior Closure Knee Wrap w/Hinge may be hand washed using warm water and detergent for delicate fabrics. Repeatedly rinse and line dry. Do not use bleach or place in a washer or dryer. Hinges are removable.

Caution

This product has been designed and tested based on single patient usage. Since it is a single use device it should never be used by multiple patients. If additional pain or symptoms occur while using this device, seek medical attention.

Indications

Medial/lateral instability, muscle/ligament weakness or injury, degenerative joint disease, post-op rehabilitation, and arthritis.

Contraindications

Not for use on patients with circulatory or sensation issues. If swelling, pain, skin irritation, or an unusual reaction occurs, discontinue use and consult your medical professional.

Warranty

Products manufactured by Corflex are warranted to the original purchaser against defects in workmanship and material for the life of the product. The foregoing warranty shall not apply to normal wear and tear; or if the product has been subjected to misuse or misapplication the warranty will be voided at the discretion of Corflex Global.

To Reorder

Product Number		Size	Measurement		
Open Popiteal	Closed Popiteal		6" Above Mid-Pat	Mid-Pat	6" Below Mid-Pat
86-1302	—	XS	14"-16"	10"-12"	9"-11"
86-1303	86-1313	S	15"-17"	12"-14"	11"-13"
86-1304	86-1314	M	17"-19"	14"-16"	13"-14"
86-1305	86-1315	L	19"-21"	16"-18"	14"-16"
86-1306	86-1316	XL	21"-23"	18"-19"	16"-18"
86-1307	86-1317	2XL	22"-24"	19"-21"	18"-20"
86-1308	86-1318	3XL	23"-26"	21"-23"	20"-22"
86-1309	86-1319	4XL	26"-29"	23"-26"	22"-24"



669 EAST INDUSTRIAL PARK DRIVE
MANCHESTER, NH 03109
CORFLEXGLOBAL.COM | 800 426 7353

Care Instructions

The CoolTex Knee Anterior Closure Knee Wrap w/Hinge may be hand washed using warm water and detergent for delicate fabrics. Repeatedly rinse and line dry. Do not use bleach or place in a washer or dryer. Hinges are removable.

Caution

This product has been designed and tested based on single patient usage. Since it is a single use device it should never be used by multiple patients. If additional pain or symptoms occur while using this device, seek medical attention.

Indications

Medial/lateral instability, muscle/ligament weakness or injury, degenerative joint disease, post-op rehabilitation, and arthritis.

Contraindications

Not for use on patients with circulatory or sensation issues. If swelling, pain, skin irritation, or an unusual reaction occurs, discontinue use and consult your medical professional.

Warranty

Products manufactured by Corflex are warranted to the original purchaser against defects in workmanship and material for the life of the product. The foregoing warranty shall not apply to normal wear and tear; or if the product has been subjected to misuse or misapplication the warranty will be voided at the discretion of Corflex Global.

To Reorder

Product Number		Size	Measurement		
Open Popiteal	Closed Popiteal		6" Above Mid-Pat	Mid-Pat	6" Below Mid-Pat
86-1302	—	XS	14"-16"	10"-12"	9"-11"
86-1303	86-1313	S	15"-17"	12"-14"	11"-13"
86-1304	86-1314	M	17"-19"	14"-16"	13"-14"
86-1305	86-1315	L	19"-21"	16"-18"	14"-16"
86-1306	86-1316	XL	21"-23"	18"-19"	16"-18"
86-1307	86-1317	2XL	22"-24"	19"-21"	18"-20"
86-1308	86-1318	3XL	23"-26"	21"-23"	20"-22"
86-1309	86-1319	4XL	26"-29"	23"-26"	22"-24"



669 EAST INDUSTRIAL PARK DRIVE
MANCHESTER, NH 03109
CORFLEXGLOBAL.COM | 800 426 7353