

CORFLEX®


Designed for life.

CORFLEXGLOBAL.COM | 800 426 7353


Marathon Speed Lacer™ Ankle Brace Application Instructions

Fitting and Application:


1. Plastic stabilizers are included with each brace and located in a vertical pocket on each side of the ankle. Release the loop tab at the top of the pocket and remove if desired. Brace is designed to be worn over an athletic sock.





2. Slide brace on foot and position your heel into the hole at the bottom of the brace. Wrap ankle laces forward across the top of foot and secure tab on hook and loop. Be sure that lace tabs are positioned so they fall outside of shoe area to prevent discomfort.



3. Grasp medial (inside) figure-8 stabilizing strap and wrap forward across the top of the foot. Continue wrapping strap under the heel and attach hook end of strap to the side of the boot portion of the brace. Repeat process for the lateral (outside) strap.



4. Tighten the figure-8 stabilizing straps from a seated position with your foot flat on the floor. Use the finger loops to first free the straps from the loop on the side of the boot. Next pull up firmly on both straps, then press straps back against boot securely.



CORFLEX®


Designed for life.

CORFLEXGLOBAL.COM | 800 426 7353


Marathon Speed Lacer™ Ankle Brace Application Instructions

Fitting and Application:


1. Plastic stabilizers are included with each brace and located in a vertical pocket on each side of the ankle. Release the loop tab at the top of the pocket and remove if desired. Brace is designed to be worn over an athletic sock.




2. Slide brace on foot and position your heel into the hole at the bottom of the brace. Wrap ankle laces forward across the top of foot and secure tab on hook and loop. Be sure that lace tabs are positioned so they fall outside of shoe area to prevent discomfort.



3. Grasp medial (inside) figure-8 stabilizing strap and wrap forward across the top of the foot. Continue wrapping strap under the heel and attach hook end of strap to the side of the boot portion of the brace. Repeat process for the lateral (outside) strap.




4. Tighten the figure-8 stabilizing straps from a seated position with your foot flat on the floor. Use the finger loops to first free the straps from the loop on the side of the boot. Next pull up firmly on both straps, then press straps back against boot securely.



APPLICATION INSTRUCTIONS (continued)

5. Wrap 3" elastic circumferential strap over tied laces and secure strap.



Care Instructions

The Marathon Speed Lacer Ankle Stabilizer may be hand washed using warm water and detergent for delicate fabrics. Remove plastic stays prior to washing. Repeatedly rinse and line dry. Do not use bleach or place in a washer or dryer. Plastic stays can be cleaned using warm water and a mild disinfectant.

Caution

This product has been designed and tested based on single patient use. Since it is a single use device it should never be used by multiple patients. If any additional pain or symptoms occur while using this device, seek medical attention.

Indications

Indicated for chronic and acute ankle sprains; control for abnormal ankle eversion or inversion; and as protective sportswear.

Contraindications

Do not use on fractures. Not for use on patients with circulatory or sensation issues. If swelling, pain, skin irritation or an unusual reaction occurs, discontinue use and consult your medical professional.

Warranty

Products manufactured by Corflex are warranted to the original purchaser against defects in workmanship and material for the life of the product. The foregoing warranty shall not apply to normal wear and tear; or if the product has been subjected to misuse or misapplication the warranty will be voided at the discretion of Corflex.

To Reorder

2XS	XS	S	M	L	XL	2XL	3XL
75-8130	75-8131	75-8132	75-8133	75-8134	75-8135	75-8136	75-8137

Sizing Guide - Circumference

9"-10"	10"-11"	11"-12"	12"-13"	13"-14"	14"-15"	15"-16"	16"+
--------	---------	---------	---------	---------	---------	---------	------

Measurement is circumference from base of heel around ankle.

Sizing Guide - Shoe Size


Men	6-7	7.5-9	9.5-11	11.5-13	13.5-15	15+
Narrow	2XS	XS	S	M	L	XL
Regular	XS	S	M	L	XL	2XL
Wide	S	M	L	XL	2XL	3XL



669 EAST INDUSTRIAL PARK DRIVE
MANCHESTER, NH 03109
CORFLEXGLOBAL.COM | 800 426 7353

APPLICATION INSTRUCTIONS (continued)

5. Wrap 3" elastic circumferential strap over tied laces and secure strap.



Care Instructions

The Marathon Speed Lacer Ankle Stabilizer may be hand washed using warm water and detergent for delicate fabrics. Remove plastic stays prior to washing. Repeatedly rinse and line dry. Do not use bleach or place in a washer or dryer. Plastic stays can be cleaned using warm water and a mild disinfectant.

Caution

This product has been designed and tested based on single patient use. Since it is a single use device it should never be used by multiple patients. If any additional pain or symptoms occur while using this device, seek medical attention.

Indications

Indicated for chronic and acute ankle sprains; control for abnormal ankle eversion or inversion; and as protective sportswear.

Contraindications

Do not use on fractures. Not for use on patients with circulatory or sensation issues. If swelling, pain, skin irritation or an unusual reaction occurs, discontinue use and consult your medical professional.

Warranty

Products manufactured by Corflex are warranted to the original purchaser against defects in workmanship and material for the life of the product. The foregoing warranty shall not apply to normal wear and tear; or if the product has been subjected to misuse or misapplication the warranty will be voided at the discretion of Corflex.

To Reorder

Product Number

2XS	XS	S	M	L	XL	2XL	3XL
75-8130	75-8131	75-8132	75-8133	75-8134	75-8135	75-8136	75-8137

Sizing Guide - Circumference

9"-10"	10"-11"	11"-12"	12"-13"	13"-14"	14"-15"	15"-16"	16"+
--------	---------	---------	---------	---------	---------	---------	------

Measurement is circumference from base of heel around ankle.

Sizing Guide - Shoe Size

Men	6-7	7.5-9	9.5-11	11.5-13	13.5-15	15+
Narrow	2XS	XS	S	M	L	XL
Regular	XS	S	M	L	XL	2XL
Wide	S	M	L	XL	2XL	3XL



669 EAST INDUSTRIAL PARK DRIVE
MANCHESTER, NH 03109
CORFLEXGLOBAL.COM | 800 426 7353